



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

WiScale, LLC dba UniFide CST
4123 Terminal Drive
McFarland, WI 53558

Fulfills the requirements of

ISO/IEC 17025:2017

and the national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 07 September 2027

Certificate Number: AC-3371.01



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

WiScale, LLC dba UniFide CST

4123 Terminal Dr.
McFarland, WI 53558
Dan Christianson 800-747-4474

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **03 September 2025**

Certificate Number: **AC-3371.01** Certificate Expiry Date: **07 September 2027**

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
5-Place Balances ¹ (10, 20, & 50 µg resolution)	Up to 100 g	1d + 0.000 15% load	Comparison to ASTM E617 Class 0 Weights
4-Place Balances ¹ (0.1, 0.2 & 0.5 mg resolution)	Up to 400 g	1d + 0.000 13% load	Comparison to ASTM E617 Class 0 Weights
4-Place Balances ¹ (0.1, 0.2 & 0.5 mg resolution)	Up to 1 000 g	1d + 0.000 30% load	Comparison to ASTM E617 Class 1 Weights
Class I/Equivalent Balances ¹ (1 mg and greater resolution)	Up to 15 000 g	1d + 0.000 30% load	Comparison to ASTM E617 Class 1 Weights
Class II/Equivalent Balances ¹ (1 mg and greater resolution)	Up to 15 000 g	0.6d + 0.000 065% load	
Class II/Equivalent Balances ¹ (1 mg and greater resolution)	Up to 80 000 g	0.6d + 0.000 25% load	Comparison to ASTM E617 Class 1 & 2 Weights
Class III/Equivalent Scales ¹	Up to 50 000 lb Up to 20 000 kg	1d + 0.004 0% load	Comparison to NIST Class F Weights

This Scope of Accreditation, version 002, was last updated on: 03 September 2025 and is valid only when accompanied by the Certificate.

Page 1 of 2

1899 L Street NW, Suite 1100-A, Washington, DC 20036

414-501-5494

www.anab.org



Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-) ²	Reference Standard, Method, and/or Equipment
Class IIIIL Vehicle Scales ¹	Up to 200 000 lb	1d + 0.004 0% load	Comparison to NIST Class F Weights
High-Resolution Unmarked Scales ¹	Up to 50 000 lb	1d + 0.017% load	Comparison to NIST Class F Weights
High-Resolution Unmarked Scales ¹	Up to 15 000 g	1d + 0.000 30% load	Comparison to ASTM E617 Class 1 Weights
	Up to 80 000 g	0.6d + 0.000 25% load	Comparison to ASTM E617 Class 1 & 2 Weights
	Up to 100 000 g	1d + 0.001 1% load	Comparison to ASTM E617 Class 1, 2, & 3 Weights
	Up to 20 000 kg	1d + 0.012% load	Comparison to NIST Class F Weights

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.



Jason Stine, Vice President