



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

**WiScale, LLC dba UniFide CST**  
4123 Terminal Drive  
McFarland, WI 53558

Fulfills the requirements of

**ISO/IEC 17025:2017**

and the national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

**CALIBRATION**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 07 September 2027  
Certificate Number: AC-3371.01



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**AND**

**ANSI/NCSL Z540-1-1994 (R2002)**

**WiScale, LLC dba UniFide CST**

4123 Terminal Dr.  
McFarland, WI 53558

Brook Whitman 608-731-2652

**CALIBRATION**

ISO/IEC 17025 Accreditation Granted: **03 September 2025**

Certificate Number: **AC-3371.01** Certificate Expiry Date: **07 September 2027**

**Mass and Mass Related**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
5-Place Balances <sup>1</sup> (10, 20, & 50 µg resolution)	Up to 100 g	1d + 0.000 15% load	ASTM E617 Class 0 Weights
4-Place Balances <sup>1</sup> (0.1, 0.2 & 0.5 mg resolution)	Up to 400 g	1d + 0.000 13% load	ASTM E617 Class 0 Weights
4-Place Balances <sup>1</sup> (0.1, 0.2 & 0.5 mg resolution)	Up to 1 000 g	1d + 0.000 30% load	ASTM E617 Class 1 Weights
Class I/Equivalent Balances <sup>1</sup> (1 mg and greater resolution)	Up to 15 000 g	1d + 0.000 30% load	ASTM E617 Class 1 Weights
Class II/Equivalent Balances <sup>1</sup> (1 mg and greater resolution)	Up to 15 000 g	0.6d + 0.000 065% load	ASTM E617 Class 1 Weights
Class II/Equivalent Balances <sup>1</sup> (1 mg and greater resolution)	Up to 80 000 g	0.6d + 0.000 25% load	ASTM E617 Class 1 & 2 Weights


**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Class III/Equivalent Scales <sup>1</sup>	Up to 50 000 lb Up to 20 000 kg	1d + 0.004 0% load	NIST Class F Weights
Class IIIIL Vehicle Scales <sup>1</sup>	Up to 200 000 lb	1d + 0.004 0% load	NIST Class F Weights
High-Resolution Unmarked Scales <sup>1</sup>	Up to 50 000 lb	1d + 0.017% load	NIST Class F Weights
High-Resolution Unmarked Scales <sup>1</sup>	Up to 15 000 g	1d + 0.000 30% load	ASTM E617 Class 1 Weights
	Up to 80 000 g	0.6d + 0.000 25% load	ASTM E617 Class 1 & 2 Weights
	Up to 100 000 g	1d + 0.001 1% load	ASTM E617 Class 1, 2, & 3 Weights
	Up to 20 000 kg	1d + 0.012% load	NIST Class F Weights

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. d = least resolution digit
3. WiScale has resident technicians in the Madison and Milwaukee areas.



Jason Stine, Vice President